

***** **CERTIFICATE OF ANALYSIS** *****

Article: COMPOFLOW®4F T&B-63 ML CPD/100 ML SAG-M-RCC+PDS-V
 Article number: CQ32250
 Anticoagulant solution: 63 ml CPD
 Preservation solution: 100 ml SAG-Mannitol
 Batch number: 41YD09FA00
 Manufacturing date: 9 April 2026
 Quantity: 23 472
 Expiry date: 2028-03

Test results

1. CPD Anticoagulant solution

	Method¹	Nominal value	Limits	Results
Sodium citrate.2H ₂ O	HPLC	26,3 g/l	25,0 – 27,6 g/l	25,4 g/l
Citric acid.1H ₂ O	HPLC	3,27 g/l	2,94 – 3,60 g/l	3,20 g/l
NaH ₂ PO ₄ .2H ₂ O	HPLC or Spectr.Ph.	2,51 g/l	2,38 – 2,64 g/l	2,42 g/l
Glucose.1H ₂ O ²	HPLC	25,5 g/l	24,2 – 26,8 g/l	24,6 g/l
5-HMF	HPLC		< 5 ppm	0,4 ppm
pH	Potent.	5,6	5,3 – 5,9	5,6
Volume	Volumetr.	63 ml	57 – 69 ml	66 ml


2. SAG-Mannitol Preservation solution

	Method	Nominal value	Limits	Results
Sodium chloride	Argent.	8,77 g/l	8,33 – 9,21 g/l	8,63 g/l
Adenine	HPLC	0,169 g/l	0,161 – 0,177 g/l	0,166 g/l
Glucose.1H ₂ O ²	HPLC	9,00 g/l	8,55 – 9,45 g/l	8,77 g/l
Mannitol	HPLC	5,25 g/l	4,99 – 5,51 g/l	5,09 g/l
5-HMF	HPLC		< 5 ppm	0,4 ppm
pH	Potent.	5	3,5 – 6,0	5,2
Volume	Volumetr.	100 ml	90 – 110 ml	104 ml

¹⁾ The used test methods are equivalent or more accurate than the corresponding methods described in the Eur. Pharm.

²⁾ As total reducing sugars

3. Characteristics: clear, colourless to faintly yellow solution practically free from visible particles conform
4. Physical control: in accordance with specifications conform
5. Sterility (Eur. Pharm.): sterile conform
6. Bacterial Endotoxines (Eur. Pharm.):
 CPD: < 5,56 EU/ml conform
 SAGM: < 5,56 EU/ml conform
7. Used Filterbatch 26GY001,26GY002,26GY003,26RF010

<u>Batch released by Quality Assurance.</u>		<u>Manufactured by:</u>	
Date:	28.04.2026	Fresenius Kabi AG	
Name:	Klímová Vladimíra	Plant: Fresenius Kabi Horátev CZ s.r.o.	
Signature:		Hořátev 104, 289 13 Czech Republic	