

******* CERTIFICATE OF ANALYSIS *******

Article: CompoStop Flex2F PLT processing and leuko red.syst

Article number: FD51600

Batch number: 41XE05GA00 Quantity: 15 480

Manufacturing date: 7 May 2025 Expiry date: 2027-04

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA006	

<u>Batch released by Quality Assurance.</u>		<u>Manufactured by:</u>	
Date:	28.05.2025	Fresenius Kabi AG	
Name:	Tykva František	Plant:	
Signature:		Fresenius Kabi Horátev CZ s.r.o.	
		Hořátev 104, 289 13	
		Czech Republic	