



CERTIFICATE OF ANALYSIS

Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41XG23GA00	Quantity:	15 400
Manufacturing date:	24 July 2025	Expiry date:	2027-06

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA014	

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 15.08.2025	Fresenius Kabi AG
Name: Tykva František	Plant:
Signature: 	Fresenius Kabi Horatev CZ s.r.o.
	Hořátev 104, 289 13
	Czech Republic

Form.no. 17-4269/2