



Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41XK28GA00	Quantity:	15 480
Manufacturing date:	31 October 2025	Expiry date:	2027-09

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA024,25PA025	

<u>Batch released by Quality Assurance.</u> Date: 20.11.2025 Name: Tykva František Signature: 	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi Horátev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
--	---

Form.no. 17-4269/2