



**FRESENIUS
KABI**

***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst
Article number: FD51600
Batch number: 41XM15GA00 Quantity: 5 920
Manufacturing date: 28 January 2026 Expiry date: 2027-11

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA037	

<u>Batch released by Quality Assurance.</u> Date: 13.02.2026 Name: Tykva František Signature: 	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi Horátev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
--	---

Form.no. 17-4269/2