



**FRESENIUS  
KABI**

\*\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*\*

Article: CompoStop Flex2F PLT processing and leuko red.syst  
Article number: FD51600  
Batch number: 41YB03GA00 Quantity: 9 520  
Manufacturing date: 16 February 2026 Expiry date: 2028-01

<b>Test</b>	<b>Limits</b>	<b>Results</b>
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	26PA001	

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 05.03.2026	Fresenius Kabi AG
Name: Tykva František	Plant:
Signature: 	Fresenius Kabi Horatev CZ s.r.o.
	Hořátev 104, 289 13
	Czech Republic

Form.no. 17-4269/2