



**FRESENIUS
KABI**

***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst
Article number: FD51600
Batch number: 41YB17GA00 Quantity: 15 480
Manufacturing date: 13 February 2026 Expiry date: 2028-01

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	26PA001,26PA002	

<u>Batch released by Quality Assurance.</u>	<u>Manufactured by:</u>
Date: 03.03.2026	Fresenius Kabi AG
Name: Tykva František	Plant: Fresenius Kabi Horatev CZ s.r.o.
Signature: 	Hořátev 104, 289 13 Czech Republic

Form.no. 17-4269/2