FRESENIUS KABI

***** **CERTIFICATE OF ANALYSIS** *****

Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41YD08GA00	Quantity:	15 480
Manufacturing date:	15 April 2026	Expiry date:	2028-03

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	26PA007,26PA008	

Batch released by Quality Assurance.		Manufactured by:
Date:	07.05.2026	Fresenius Kabi AG Plant: Fresenius Kabi Horatev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
Name:	Klímová Vladimíra	
Signature:		

Form.no. 17-4269/2