



***** **CERTIFICATE OF ANALYSIS** *****

Article: CompoStop Flex2F PLT processing and leuko red.syst

Article number: FD51600

Batch number: 41YE18GA00 Quantity: 15 440

Manufacturing date: 21 May 2026 Expiry date: 2028-04

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	26PA011,26PA012	

<u>Batch released by Quality Assurance.</u>		<u>Manufactured by:</u>
Date:	16.06.2026	Fresenius Kabi AG
Name:	Tykva František	Plant:
Signature:		Fresenius Kabi Horátev CZ s.r.o.
		Hořátev 104, 289 13
		Czech Republic