FRESENIUS KABI

***** CERTIFICATE OF ANALYSIS *****

Article:	CompoStop Flex2F PLT processing and leuko red.syst		
Article number:	FD51600		
Batch number:	41XH30GA00	Quantity:	14 520
Manufacturing date:	30 August 2025	Expiry date:	2027-07

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform
Used Filterbatch	25PA018,25PA021	

Batch released by Quality Assurance.		Manufactured by:
Date:	19.09.2025	Fresenius Kabi AG Plant: Fresenius Kabi Horatev CZ s.r.o. Hořátev 104, 289 13 Czech Republic
Name:	Tykva František	
Signature:		

Form.no. 17-4269/2