FRESENIUS KABI

**** CERTIFICATE OF ANALYSIS ****

Article: CompoStop 2F – PLT processing and leuko reduction system
Article number: ND51600

Batch Number: 41YG29GA00
Manufacturing Date: 21-Jul-2026

Quantity: 15480 pieces
Expiration Date: Jun-2028

	Limits	Results
Configuration	Conform product specification	conform
Physical control	In accordance with specifications	conform
Bacterial Endotoxines	< 20 EU/device	conform
Sterility	Sterile (Ph.Eur.)	conform
Used Filter batch		26PA018, 26PA024

Batch released by Quality Assurance.

Date: 19-Aug-2026
Name: František Tykva
Signature:

Manufactured by:
Fresenius Kabi
Plant: Fresenius Kabi Horatev CZ s.r.o.
Hořátev 104
289 13
Czech Republic

Form.no. 17-4269/2

Secure Report ID: 2234
LIMS Lot Num: 2756
COA ID: 1,991

CoA Setup: Finished Product_EN v1
Print Date: 19.08.2026 12:41
Database: LIMS_PROD