

\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*

Article:	Compoflex® 1F Empty - PD-V	
Article number:	T2131	
Batch number:	85WG12CA00	Quantity: 5543
Manufacturing date:	2024.07	Expiry date: 2027.06

<b>Test</b>	<b>Limits</b>	<b>Results</b>
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 02 Aug. 2024 Name: Wei Yuchan Signature: Wei Yuchan	<u>Manufactured by:</u> Fresenius Kabi AG Plant: FKGZ
--	---