

**** **CERTIFICATE OF ANALYSIS** ****

Article: CompoFlex® 1F Empty - PD-V
Article number: T2131
Batch number: 85WH08CA00 Quantity: 2761
Manufacturing date: 2024.08 Expiry date: 2027.07

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 26 Aug. 2024 Name: Wei Yuchan Signature: Wei Yuchan	<u>Manufactured by:</u> Fresenius Kabi AG Plant: FKGZ
--	---