

\*\*\*\* **CERTIFICATE OF ANALYSIS** \*\*\*\*

Article: CompoFlex® 1F Empty - PD-V  
Article number: T2131  
Batch number: 85XD14CA00 Quantity: 9141  
Manufacturing date: 2025.04 Expiry date: 2028.03

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 30 Apr. 2025 Name: Wei Yuchan Signature: Wei Yuchan	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi (Guangzhou) Co., Ltd.
--	---