

****** CERTIFICATE OF ANALYSIS ******

Article: CompoFlex® 1F Empty - PD-V
Article number: T2131
Batch number: 85XE08CB00 Quantity: 11781
Manufacturing date: 2025.05 Expiry date: 2028.04

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

Batch released by Quality Assurance.

Date: 26 May 2025

Name: Wei Yuchan

Signature: Wei Yuchan

Manufactured by:

Fresenius Kabi AG

Plant: Fresenius Kabi (Guangzhou) Co., Ltd.