

****** CERTIFICATE OF ANALYSIS ******

Article:	Compoflex® 1F Empty - PD-V		
Article number:	T2131		
Batch number:	85XI18CA00	Quantity:	10307
Manufacturing date:	2025.09	Expiry date:	2028.08

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 21 Oct 2025 Name: Xu Xiaoying Signature: <i>Xu Xiaoying</i>	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi (Guangzhou) Co., Ltd.
--	---