

****** CERTIFICATE OF ANALYSIS ******

Article: CompoFlex® 1F Empty - PD-V
Article number: T2131
Batch number: 85YC12CC00 Quantity: 2750
Manufacturing date: 2026.03 Expiry date: 2029.02

Test	Limits	Results
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

<u>Batch released by Quality Assurance.</u> Date: 27 Mar. 2026 Name: Xu Xiaoying Signature: 	<u>Manufactured by:</u> Fresenius Kabi AG Plant: Fresenius Kabi (Guangzhou) Co., Ltd.
--	---