FRESENIUS KABI

**** CERTIFICATE OF ANALYSIS ****

Article:	Compoflex® 1F Empty - PD-V		
Article number:	T2131		
Batch number:	85YD02CD00	Quantity:	7340
Manufacturing date:	2026.04	Expiry date:	2029.03

Test	**Limits**	**Results**
Configuration	Conform product specification	Conform
Physical control	In accordance with specifications	Conform
Bacterial Endotoxines	< 20 EU/device	Conform
Sterility	Sterile (Ph.Eur.)	Conform

Batch released by Quality Assurance.	Manufactured by:
Date: 21 Apr. 2026	Fresenius Kabi AG
Name: Xu Xiaoying	Plant: Fresenius Kabi (Guangzhou) Co., Ltd.
Signature: Xu Xiaoying	

Form.no. 17-4228/2