



## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
(Class IIa and Class IIb Devices)

**No. G10 037258 0014 Rev. 05**

### Manufacturer:

**Fresenius Kabi AG**

Else-Kröner-Str. 1  
61352 Bad Homburg  
GERMANY

SRN Manufacturer - DE-MF-000009273

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result.

The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with.

For details and certificate validity see: [www.tuvsud.com/ps-cert?q=cert:G10 037258 0014 Rev. 05](http://www.tuvsud.com/ps-cert?q=cert:G10_037258_0014_Rev_05)

**Report No.:** 713235231 / 713310265

**Preceding Certificate No.:** G10 037258 0014 Rev. 04

**Valid from:** 2024-09-26

**Valid until:** 2026-10-17

**Date of Initial Issuance:** 2021-10-18

Christoph Dicks  
Head of Certification/Notified Body

**Issue date:** 2024-09-26



## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
(Class IIa and Class IIb Devices)

**No. G10 037258 0014 Rev. 05**

<b>Classification:</b>	Class IIb
<b>Device Group:</b>	B020101 - BED-SIDE LEUKOREDUCTION FILTERS
<b>Intended Purpose:</b>	Filters are medical devices intended for leukocyte reduction of blood components for transfusion.
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	Z121704 - APHERESIS EQUIPMENT
<b>Intended Purpose:</b>	Apheresis Devices intended for the collection of blood components. Apheresis Devices intended for the collection of blood components and/or therapeutic apheresis.
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	B020102 - LABORATORY LEUKOREDUCTION FILTERS
<b>Intended Purpose:</b>	Filters are medical devices intended for leukocyte reduction of blood components for transfusion.
<b>Classification:</b>	Class IIa
<b>Device Group:</b>	Z121780 - BLOOD TRANSFUSION INSTRUMENTS - HARDWARE ACCESSORIES
<b>Intended Purpose:</b>	-/-
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	B0401 - INTRA- AND POSTOPERATIVE BLOOD COLLECTION, WASHING AND REINFUSION DEVICES AND KITS
<b>Intended Purpose:</b>	Processing of autologous shed blood collected intra-operatively and postoperatively to obtain washed packed red blood cells for reinfusion or for peri-operative separation of blood into Packed Red Cells, Plasma and Platelet Rich Plasma.
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	B030204 - EXTRACORPOREAL PHOTOCHAEMOTHERAPY OR PHOTOPHERESIS DEVICES AND KITS
<b>Intended Purpose:</b>	Apheresis devices intended for the administration of Extracorporeal Photopheresis (ECP).
<b>Classification:</b>	Class IIb
<b>Device Group:</b>	B030202 - CYTAPHERESIS DEVICES AND KITS
<b>Intended Purpose:</b>	Apheresis Devices intended for the collection of blood components. Apheresis devices intended for the collection of blood components and/or therapeutic apheresis



## EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III  
 (Class IIa and Class IIb Devices)

**No. G10 037258 0014 Rev. 05**

**Classification:** Class IIb  
**Device Group:** B030201 - PLASMAPHERESIS DEVICES AND KITS  
**Intended Purpose:** Apheresis devices intended for the collection of blood components.

**Classification:** Class IIb  
**Device Group:** B010299 - BLOOD TRANSFER BAGS AND KITS - OTHER  
**Intended Purpose:** Blood Transfer Bags are for the processing, filtration and/or storage of whole blood or blood components

**The validity of this certificate depends on conditions and/or is limited to the following:** none

### Revision History:

Rev.	Dated	Report	Description
00	2021-10-18	713198383	-
01	2022-03-10	713215445	-
02	2022-10-25	713235232	-
03	2022-11-18	713235235	-
04	2023-06-21	713207465 / 713211642	Supplemented: Device(s)/group of device(s) added
05	2024-09-26	713235231 / 713310265	Supplemented: Device(s)/group of device(s) added